IHE Integration Statement					
Vendor	Product Name	Version	Date		
Canon Medical Systems Corporation	Xario [™] 100 TUS-X100S	V4.0	9 September. 2016		
Integration Profiles Implemented	Actors Implemented	Options Implemented			
Radiology	IHE Radiology Technical Framewor	k (Rev 12.0)			
Scheduled Workflow	Acquisition Modality	Patient Based Worklist Query Broad Worklist Query PPS Exception Management Assisted Acquisition Protocol Setting Modality Group Case			
Patient Information Reconciliation	Acquisition Modality	No options d	No options defined		
Portable Data for Imaging	Portable Media Creator	None	None		
Evidence Documents	Acquisition Modality	No options defined TID 5000 OB-GYN Ultrasound Procedure Report			
Cardiology	IHE Cardiology Technical Framewo	ork (Rev 5.0)			
Echocardiography Workflow	Acquisition Modality	Patient Based Worklist Query Broad Worklist Query PPS Exception Management Stress Echo			
Evidence Documents	Acquisition Modality	TID 5100 \ Ultrasound TID 5200 E	Echo Evidence TID 5100 Vascular Ultrasound Report TID 5200 Echocardiography Procedure Report		
Stress Testing Workflow	Acquisition Modality	Broad Workli	d Worklist Query ist Query on Management		
IT Infrastructure	IHE IT Infrastructure Technical Fra	mework (Rev 10).1)		
Consistent Time	Time Client	None			
Internet address for vendor Japan: https://www.medical.canon/ Global: https://www.medical.canon/	Interoperability/IHE/JP	•			

Links to Standards Conformance Statements for the Implementation				
HL7	None			
DICOM	Japan: https://www.medical.canon/Interoperability/DICOM/JP Global: https://www.medical.canon/Interoperability/DICOM/EN			
Links to general information on IHE				
In North America: http://www.ihe.net		In Europe: http://www.ihe-europe.net	In Japan: http://www.ihe-j.org	

[&]quot;Xario" is a trademark of Canon Medical Systems Corporation.