

IHE Integration Statement			
Vendor	Product Name	Version	Date
Canon Medical Systems Corporation	Aquilion ONE TSX-308A	V1.3 or later	25 Sep. 2023
Integration Profiles Implemented	Actors Implemented	Options Implemented	
Radiology IHE Radiology Technical Framework			
Scheduled Workflow.b	Acquisition Modality	Patient Based Worklist Query Broad Worklist Query Assisted Acquisition Protocol Setting Modality Group Case PPS Exception Management Billing and Material Management Japanese Character Set	
Patient Information Reconciliation	Acquisition Modality	No options defined	
Consistent Presentation of Images	Acquisition Modality	No options defined	
Charge Posting	Acquisition Modality	PPS Exception Management Modality Group Case	
Presentation of Grouped Procedures	Acquisition Modality	No options defined	
Portable Data for Imaging	Portable Media Creator	None	
Radiation Exposure Monitoring	Acquisition Modality	No options defined	
IT Infrastructure IHE IT Infrastructure Technical Framework			
Audit Trail and Node Authentication	Secure Application Secure Node	STX: TLS 1.2 Floor using BCP195 ATX: TLS Syslog ATX: UDP Syslog	
<u>Internet address for vendor's IHE information:</u> Japan: https://www.medical.canon/Interoperability/IHE/JP Global: https://www.medical.canon/Interoperability/IHE/EN			

Links to Standards Conformance Statements for the Implementation		
HL7	None	
DICOM	Japan: https://www.medical.canon/Interoperability/DICOM/JP Global: https://www.medical.canon/Interoperability/DICOM/EN	
Links to general information on IHE		
In North America: http://www.ihe.net	In Europe: http://www.ihe-europe.net	In Japan: http://www.ihe-j.org

HL7® is the registered trademark of Health Level Seven International and the use of this trademark does not constitute an endorsement by HL7.

DICOM® is the registered trademark of the National Electrical Manufacturers Association for its Standards publications relating to digital communications of medical information.