<table>
<tr><td colspan="4" align="center">

IHE Integration Statement

</td></tr>
<tr><th>Vendor</th><th>Product Name</th><th>Version</th><th>Date</th></tr>
<tr><td>Canon Medical Systems Corporation</td><td>Xario™ 200 TUS-X200
Xario™ 100 TUS-X100</td><td>V2.0</td><td>February 2014</td></tr>
<tr><td colspan="4">This product implements all transactions required in the IHE Radiology Technical Framework (Rev 12.0), IHE Cardiology Technical Framework (Rev 5.0) and IHE IT Infrastructure Technical Framework (Rev 10.1) to support the IHE Integration Profiles, Actors and Options listed below:</td></tr>
<tr><td colspan="2" align="center">

Integration Profiles Implemented

</td><td align="center">

Actors Implemented

</td><td align="center">

Options Implemented

</td></tr>
<tr><td colspan="4" align="center">

Radiology

</td></tr>
<tr><td colspan="2">Scheduled Workflow</td><td>Acquisition Modality</td><td>Patient Based Worklist Query
Broad Worklist Query
PPS Exception Management
Assisted Acquisition Protocol Setting
Modality Group Case</td></tr>
<tr><td colspan="2">Patient Information Reconciliation</td><td>Acquisition Modality</td><td>No options defined</td></tr>
<tr><td colspan="2">Portable Data for Imaging</td><td>Portable Media Creator</td><td>None</td></tr>
<tr><td colspan="2">Evidence Documents</td><td>Acquisition Modality</td><td>No options defined
  TID 5000 OB-GYN Ultrasound Procedure Report</td></tr>
<tr><td colspan="4" align="center">

Cardiology

</td></tr>
<tr><td colspan="2">Echocardiography Workflow</td><td>Acquisition Modality</td><td>Patient Based Worklist Query
Broad Worklist Query
PPS Exception Management
Stress Echo</td></tr>
<tr><td colspan="2">Evidence Documents</td><td>Acquisition Modality</td><td>Echo Evidence
  TID 5100 Vascular Ultrasound Report
  TID 5200 Echocardiography Procedure Report</td></tr>
<tr><td colspan="2">Stress Testing Workflow</td><td>Acquisition Modality</td><td>Patient Based Worklist Query
Broad Worklist Query
PPS Exception Management
Stress Echo</td></tr>
<tr><td colspan="4" align="center">

IT Infrastructure

</td></tr>
<tr><td colspan="2">Consistent Time</td><td>Time Client</td><td>None</td></tr>
<tr><td colspan="4">

Internet address for vendor's IHE information:

Japan:
  https://www.medical.canon/Interoperability/IHE/JP
Global:
  https://www.medical.canon/Interoperability/IHE/EN

</td></tr>
</table>

Links to Standards Conformance Statements for the Implementation	
HL7	None
DICOM	Japan: https://www.medical.canon/Interoperability/DICOM/JP Global: https://www.medical.canon/Interoperability/DICOM/EN
Links to general information on IHE	

In North America: http://www.ihe.net/	In Europe: http://www.ihe-europe.net/	In Japan: http://www.ihe-j.org/

"Xario" is a trademark of Canon Medical Systems Corporation.